

Introduction

Demonstration of Cardiac Muscle Antibody (CMA) by utilizing the indirect fluorescent antibody method enables serologic assessment or possible detection of cardiac disease. The presence of a (histologically defined) circulating antibody to one or more of the cardiac muscle antigens can aid in patient diagnosis and prognosis of such disease as; rheumatic fever, myocardial infarction and a variety of post-cardiotomy states. The substrate utilized in this kit is sections of monkey heart.

Principles

The presence of CMA has been reported in 25-43% of cases of active rheumatic fever. The level of antibody does decrease with remission of the active disease. Patients with many attacks of rheumatic fever are more likely to demonstrate CMA than those with relatively few attacks.

Myocardial infarction patients have been shown to demonstrate CMA (28-31%). The detection of this antibody in acute myocardial infarction cases and rather rare occurrence of the antibody in coronary insufficient cases without infarction can be useful information in a differential diagnosis between the two diseases.

Post cardiotomy patients have demonstrated CMA.

Materials Provided:

Storage & Stability of Components:

1. FITC Conjugate No. 1502L (3.0 ml) is to be stored at 2-8 C upon receipt. The conjugate is stable at this temperature until expiration date on the vial label.
2. The antigen slides of monkey heart sections must be stored at 2-8 C or lower upon receipt. Check label for specific expiration date.
3. CMA positive control No. 5402L (1.0 ml) should be stored at 2-8 C upon receipt. Check label for specific expiration date.
4. Universal negative control No. 1000L (1.0 ml) should be stored at 2-8 C or lower upon receipt. Check label for specific expiration date.
5. Buffer Pack No. 1601 - Phosphate Buffered Saline is stable at room temperature storage for 5

years. The reconstituted Buffer does not contain preservatives and should be stored at 2-8 C. Care should be taken to avoid contamination.

6. Mounting Medium No. 1610 is stable when stored at 2-8 C. Check label for specific expiration date.

Additional Materials Required but not Provided:

Test tubes and rack or microtiter system
Disposable pipettes
Staining Dish and Slide Forceps
Moisture Chamber
Volumetric Flask (500 ml)
Distilled H₂O
Fluorescence Microscope
Paper Towels - lint free

Reagent Preparation:

1. Buffer Pack No. 1601. Rehydrate buffer with 1 liter of sterile distilled water.

Specimen Collection:

Serological specimens should be collected under aseptic conditions. Hemolysis is avoided through prompt separation of the serum from the clot. Serum should be stored at 2-8 C if it is to be analyzed within a few days. Serum may be held for 3 to 6 months by storage at -20 C or lower. Lipemic and strongly hemolytic serum should be avoided. When specimens are shipped at ambient temperatures, addition of a preservative such as 0.01% (thimerosal) or 0.095% sodium azide is strongly recommended.

Test Instruction:

Screening: dilute test serums 1:20 in PBS.

Titration: set up doubling dilutions of serum starting at 1:20, 1:40, 1:80, 1:160, etc.

1. Once slides reach room temperature tear slide envelope at notch. Carefully remove the slide and avoid touching the antigen areas. The slide is now ready to use.
2. Place a drop of diluted serum (20 to 30 µl) and controls over the antigen wells.
3. Place slide with patient's serum and controls in a moist chamber for 30 minutes at room temperature (approximately 24 C).

4. Remove slide from moisture chamber and tap the slide on its side to allow the serum to run off onto a piece of paper towel. Using a wash bottle, gently rinse remaining sera from slide being careful not to aim the rinse stream directly on to the well.
5. Wash in PBS for five minutes. Repeat using fresh PBS.
6. Place a blotter on the lab table with absorbent side up. Remove slide from PBS and invert so that tissue side faces absorbent side of blotter. Line up wells to blotter holes. Place slide on top of blotter. **Do not allow tissue to dry.** Wipe back of slide with dry lint free paper towel. Apply sufficient pressure to slide while wiping to absorb buffer.
7. Deliver 1 drop (20-30 µl) of conjugate per antigen well. Repeat steps 3-6.
8. Place 4-5 drops of mounting medium on slide.
9. Apply a 22 x 70 mm coverslip. Examine the slide under a fluorescent microscope. Note: To maintain fluorescence, store mounted slide in a moisture chamber placed in a dark refrigerator.

Quality Control:

1. Positive and negative serum controls must be included in each day's testing to confirm reproducibility, sensitivity and specificity of the test procedure.
2. The negative serum control should result in little (+) or no fluorescence. If this control shows bright fluorescence, either the control, antigen, conjugate or technique may be at fault.
3. The positive serum control should result in bright 3+ to 4+ fluorescence. If this control shows little or no fluorescence, either the control, antigen, conjugate or technique may be at fault.
4. In addition to positive and negative serum controls, a PBS control should be run to establish that the conjugate is free from nonspecific staining of the antigen substrate. If the antigen shows bright fluorescence in the PBS control repeat using fresh conjugate. If the antigen still fluoresces, either the conjugate or antigen may be at fault.

Results:

1. Diffuse (D) low level staining throughout the tissue is considered non-specific and should be considered negative.
2. Sarcolemmal-subsarcolemmal (SSL) staining is considered positive.
3. Intermyofibrillar (IMF) staining is considered positive.

Staining of the SSL and/or the IMF type are both considered positive and can appear together or separately in the previously described instances.

Limitations of Procedure

1. No diagnosis should be based on a single serologic test since various host factors must be taken into consideration.
2. It has been found that some strains of streptococcus cross-react with cardia antigens.
3. Heart muscle or pericardial tissue damage due to surgery or stab wounds can produce CMA.
4. This test is for In Vitro Diagnostic Use.

Precautions

1. All human components have been tested by radioimmunoassay for (HB_sA_G) and HTLVIII/LAV by an FDA approved method and found to be negative. (Not repeatedly reactive). However, this does not assure the absence of HB_sA_G or HTLVIII/LAV. All human components should be handled with appropriate care.
2. The sodium azide (0.095%) included in the controls and conjugate is toxic if ingested.
3. Do not use components beyond their expiration date.
4. Follow the procedural instructions exactly as they appear in this insert to insure valid results.
5. For In Vitro Diagnostic Use.
6. Handle slides by the edges since direct pressure on the antigen wells may damage the antigen.
7. Once the procedure has started do not allow the antigen in the wells to dry out. This may result in

false negative test results, or unnecessary artifacts.

BIBLIOGRAPHY:

1. Van Der Geld, Anti-Heart Antibodies, in past pericardiotomy and the past myocardial-infraction syndromes. *Lance*. 2:617-621, 1964.
2. Kaplan, Autoimmunity to Heart. *Textbooks of Autoimmunopath. Grune and Stratton, Vol. II, p. 641, 1969.*
3. Anderson: Autoimmunity: Clinical and Exper. Charles C. Thomas III, Chap. II 1967.
4. Dragatakis, Autoimmune Myocarditis: A clinical entity. *CMA Journal/Vol. 120 pp. 317-321. Feb. 3, 1979.*

CARDIAC MUSCLE ANTIBODY TEST SYSTEM

For In Vitro Diagnostic Use

CATALOG NO.:

5448L	48 Test
5496L	96 Test

New Liquid "Ready for Use"

 SCIMEDX CORPORATION
400 Ford Road
Denville, NJ 07834 USA
Tel: 800.221.5598 Tel: 973.625.8822
Fax: 973.625.8796
www.scimedx.com

EC	REP
----	-----

 Emergo Europe
Molenstraat 15
2513 BH The Hague
The Netherlands



*Your Preferred
Autoimmune and
Infectious Disease
Detection Company*

400 FORD ROAD
DENVER, NEW JERSEY 07834 U.S.A.
(973) 625-8822 ♦ FAX: (973) 625-8796

CUSTOMER INQUIRIES:
(800) 221-5598

Printed in U.S.A. Rev. C 12/08/03 5448L.C□